



Test Report

No.GS20180831007 Date: 2018-09-05

Page 1 of 1

Result:

1.ISO 20126:2012 Brush

Test method: ISO 20126:2012 dentistry-manual toothbrushes-general requirements and test method-Safety requirements for manual toothbrushes

Clause	Test Item	Result
4.1	Pass-Fail criteria	P
4.2	Physical inspection	P(29.3N)
4.3	Tuft retention	P
4.4	Fatigue resistance	P
4.5	Chemical challenge	P
4.6	Handle impact strength	P
7	Marking and labeling	NR
8	Packaging	NR
Note: P=Pass; F=Fail; NA=Not Applicable NC = Not Conducted; NR = Not Requested		

Sample ID.	Test Part Description
1	

*** END OF REPORT ***



Gretesting Service Shanghai Co., Ltd.
上海量达检测技术有限公司

This report is only for reference of scientific research and education, or internal quality control.

Unless otherwise stated, the results shown in this test report refer only to the sample(s) received and tested in laboratory, and such sample(s) are retained for 30days only. The sole responsibility of Gretesting Service (Shanghai) Co., Ltd is to ensure that this job has been performed according to our Quality System and complying contract and legal conditions. This company bear only limited liability such as within 5 times of the cost of the test item and do not undertake any joint liability caused by this document. Reproduction of this document is only valid if it is done completely and under the written permission of Gretesting Service (Shanghai) Co., Ltd. If you happen to have any complaints, please do it by sending email to info@gretesting.com and referring to this report number.
10/F, 1# Building, No.2899 South Lianhua Road, Minghang District, Shanghai, China
t (86-21)64788791
f (86-21)64788792

[Http://www.gretesting.com](http://www.gretesting.com)
info@gretesting.com