



测试报告

报告编号: GSyyyyymmdd*** 日期: yyyy-mm-dd

第 1 页 共 1 页

测试结果:

室内空气

项目名称	室内空气				
采样人员	***		采样日期	****年**月**日	
分析人员	***		检测日期	****年**月**-**日	
检测依据	1.GB/T 18204.2-2014 公共场所卫生检验方法 第2部分：化学污染物 2.GB 11737-1989 居住区大气中苯、甲苯和二甲苯卫生检验标准方法 3.GB/T 18883-2002 室内空气质量标准 附录C TVOC的测定 热解吸-气相色谱法				
主要检测设备及编号	1.新世纪紫外可见分光光度计 2. 气相色谱				
采样点	检测项目	检测结果	方法检出限	GB/T18883-2002 《室内空气质量标准》	单位
客厅 1#	苯	ND	0.02	0.11	mg/m ³
	二甲苯	ND	0.1	0.20	mg/m ³
	甲苯	ND	0.05	0.20	mg/m ³
	TVOC	ND	0.0005	0.60	mg/m ³
	甲醛	ND	0.01	0.10	mg/m ³
客厅 2#	苯	ND	0.02	0.11	mg/m ³
	二甲苯	ND	0.1	0.20	mg/m ³
	甲苯	ND	0.05	0.20	mg/m ³
	TVOC	ND	0.0005	0.60	mg/m ³
	甲醛	ND	0.01	0.10	mg/m ³
主卧 3#	苯	ND	0.02	0.11	mg/m ³
	二甲苯	ND	0.1	0.20	mg/m ³
	甲苯	ND	0.05	0.20	mg/m ³
	TVOC	ND	0.0005	0.60	mg/m ³
	甲醛	ND	0.01	0.10	mg/m ³
次卧 4#	苯	ND	0.02	0.11	mg/m ³
	二甲苯	ND	0.1	0.20	mg/m ³
	甲苯	ND	0.05	0.20	mg/m ³
	TVOC	ND	0.0005	0.60	mg/m ³
	甲醛	ND	0.01	0.10	mg/m ³
备注	ND 表示检测结果小于方法检出限。				

*** 报告结束 ***



Gretesting Service Shanghai Co., Ltd.
上海量远检测技术有限公司

相关项目未取得资质认定, 仅作为科研、教学或内部质量控制之用。

Unless otherwise stated, the results shown in this test report refer only to the sample(s) received and tested in laboratory, and such sample(s) are retained for 30days only. The sole responsibility of Gretesting Service (Shanghai) Co., Ltd is to ensure that this job has been performed according to our Quality System and complying contract and legal conditions. This company bear only limited liability such as within 5 times of the cost of the test item and do not undertake any joint liability caused by this document. Reproduction of this document is only valid if it is done completely and under the written permission of Gretesting Service (Shanghai) Co., Ltd. If you happen to have any complaints, please do it by sending email to info@gretesting.com and referring to this report number.
10/F, 1# Building, No.2899 South Lianhua Road, Minghang District, Shanghai, China
上海市闵行区莲花南路 2899 号莲谷科技园 1 号楼 10 层
t (86-21)64788791
f (86-21)64788792
http://www.gretesting.com
info@gretesting.com